


Certificate of performance

Article number	3 662 050 BL	Volumetric flasks 500 ml NS 19 class A blue graduation	
Batch	07-17		
Serial number		Volume	500,0000ml
Measuringnr./Batch size	10160 / 280	Tolerance	0,2500ml
Balance	Sartorius - LA 1200 S	Average volume	500,0358 ml
Operator	4	Deviation	0,0299 ml
Date	28.02.2017	Number of measurings	15



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH
 Am Bildacker 16
 97877 Wertheim
 Germany
 +49 (0) 9342 / 9301-0
 | info@witeg.de

