


# Certificate of performance

Article number 3 669 00210 BR  
Batch 05-17  
Serial number  
Measuringnr./Batch size 10069 / 147  
Balance Sartorius - LE244S-0CE  
Operator 4  
Date 31.01.2017

## Volumetric flasks trapezoidal 3 ml NS 10 Kl. A KB

Volume 3,0000ml  
Tolerance 0,0400ml  
Average volume 3,0140 ml  
Deviation 0,0053 ml  
Number of measurings 10



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
| info@witeg.de



ISO 9001  
CERTIFIED

