

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 642 010

**Volumetric flask, 100 ml, GL 18, A, conf. cert.**


**Batch:** 43-23

**Serial:**

**Measuring Nr.** 64939

**Batch size:** 100

<b>Volume:</b>	<b>100,000 ml</b>
<b>Tolerance:</b>	$\pm 0,1000$ ml
<b>Average value:</b>	100,0051 ml
<b>Deviation:</b>	0,0034 ml
<b>Precision (CV):</b>	0,0034 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	08.11.2023
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M