

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 000 BL

Vol. flasks, 5 ml, NS 7, blue grad., A, conf. cert.


Batch: 11-20

Serial:

Measuring Nr. 36690

Batch size: 498

Volume:	5,0000 ml
Tolerance:	$\pm 0,0250$ ml
Average value:	5,0036 ml
Deviation:	0,0031 ml
Precision (CV):	0,0620 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	19.03.2020
Balance:	Sartorius - MSE224S-0CE -DU

DE-M