

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 00114 BL

Vol. flask, 10 ml, NS 14, blue grad., A, conf. cert.

Batch: 51-19


Serial:

Measuring Nr. 35298

Batch size:

500

Volume:	10,000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	10,0032 ml
Deviation:	0,0155 ml
Precision (CV):	0,1551 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

19.12.2019

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M