

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 00114

Volumetric flasks, 10 ml, NS 14, A, conf. cert.

Batch: 43-19


Serial:

Measuring Nr. 34585

Batch size:

1187

Volume:	10,000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	9,9949 ml
Deviation:	0,0087 ml
Precision (CV):	0,0871 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

18.11.2019

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M