

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 002 BL

Vol. flasks, 20 ml, NS 10, blue grad., A, conf. cert.


Batch: 28-23

Serial:

Measuring Nr. 64081

Batch size: 1799

Volume:	20,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	20,0052 ml
Deviation:	0,0049 ml
Precision (CV):	0,0246 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 17.08.2023

Balance: Sartorius -
LE244S-OCE

DE-M