

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 00214

Volumetric flask, 20 ml, NS 14, A, conf. cert.


Batch: 30-22

Serial:

Measuring Nr. 58395

Batch size: 500

Volume:	20,000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	20,0024 ml
Deviation:	0,0036 ml
Precision (CV):	0,0180 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 25.08.2022

Balance: Sartorius -
LE244S-OCE

DE-M