

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 003 BL

**Vol. flasks, 25 ml, NS 10, blue grad., A, conf. cert.**

**Batch:** 47-19


**Serial:**

**Measuring Nr.** 34682

**Batch size:**

3200

<b>Volume:</b>	<b>25,0000 ml</b>
<b>Tolerance:</b>	$\pm 0,0400$ ml
<b>Average value:</b>	24,9965 ml
<b>Deviation:</b>	0,0058 ml
<b>Precision (CV):</b>	0,0234 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	9
<b>Date:</b>	28.11.2019
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M