

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 003 USP

Volumetric flasks, 25 ml, NS 10, USP, A, conf. cert.

Batch: 36-21


Serial:

Measuring Nr. 50245

Batch size:

278

Volume:	25,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	25,0038 ml
Deviation:	0,0044 ml
Precision (CV):	0,0177 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	04.11.2021
Balance:	Sartorius - MSE224S-0CE -DU

DE-M