

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 0030

Volumetric flasks, 30 ml, NS 10, A, conf. cert.

Batch: 18-24


Serial:

Measuring Nr. 67175

Batch size:

278

| | |
|------------------------|-------------|
| Volume: | 30,0000 ml |
| Tolerance: | ± 0,0400 ml |
| Average value: | 30,0059 ml |
| Deviation: | 0,0069 ml |
| Precision (CV): | 0,0231 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

16.05.2024

Balance:

Sartorius -
LE244S-OCE

DE-M