

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 0030

Volumetric flasks, 30 ml, NS 10, A, conf. cert.


Batch: 47-24

Serial:

Measuring Nr. 71118

Batch size: 280

Volume:	30,000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	30,0011 ml
Deviation:	0,0033 ml
Precision (CV):	0,0111 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	09.12.2024
Balance:	Sartorius - MSE224S-0CE -DU

DE-M