

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 005 BL

**Volumetric flasks, NS 12, blue grad., A, conf. cert.**


**Batch:** 05-18

**Serial:**

**Measuring Nr.** 18322

**Batch size:** 3198

<b>Volume:</b>	50,0000 ml
<b>Tolerance:</b>	$\pm 0,0600$ ml
<b>Average value:</b>	50,0057 ml
<b>Deviation:</b>	0,0107 ml
<b>Precision (CV):</b>	0,0214 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 9

**Date:** 22.02.2018

**Balance:** Sartorius -  
LE244S-OCE

DE-M