

Certificate of performance

Article: 3 662 005 BL

Vol. flasks, 50 ml, NS 12, blue grad., A, conf. cert.


Batch: 36-22

Serial:

Measuring Nr. 59652

Batch size: 3074

Volume:	50,000 ml
Tolerance:	± 0,0600 ml
Average value:	50,0111 ml
Deviation:	0,0077 ml
Precision (CV):	0,0154 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

12.10.2022

Balance:

Sartorius -
MSE224S-0CE
-DU