

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 005

Vol. flasks, 50 ml, NS 12, A, conf. cert

Batch: 49-17


Serial:

Measuring Nr. 17679

Batch size:

1

Volume:	50,000 ml
Tolerance:	± 0,0600 ml
Measured value:	50,0139 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	30.01.2018
Balance:	Sartorius - LE244S-OCE

DE-M