

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 005

**Vol. flasks, 50 ml, NS 12, A, conf. cert**

**Batch:** 49-17


**Serial:**

**Measuring Nr.** 17693

**Batch size:**

1

<b>Volume:</b>	<b>50,000 ml</b>
<b>Tolerance:</b>	<b>± 0,0600 ml</b>
<b>Measured value:</b>	<b>50,0213 ml</b>



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	9
<b>Date:</b>	30.01.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M