

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 006 BL

Vol. flask, 50 ml, NS 14, blue grad., A, conf. cert.


Batch: 11-22

Serial:

Measuring Nr. 54113

Batch size: 2380

Volume:	50,0000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	49,9948 ml
Deviation:	0,0072 ml
Precision (CV):	0,0144 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 05.04.2022

Balance: Sartorius -
LE244S-OCE

DE-M