

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 006

Volumetric flask, 50 ml, NS 14, A, conf. cert.


Batch: 41-19

Serial:

Measuring Nr. 33728

Batch size: 3197

Volume:	50,0000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	50,0026 ml
Deviation:	0,0130 ml
Precision (CV):	0,0259 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 17.10.2019

Balance: Sartorius -
LE244S-OCE

DE-M