

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 010 BL

Vol. flasks, 100 ml, NS 12, blue grad., A, conf. cert.

Batch: 45-20


Serial:

Measuring Nr. 42620

Batch size:

3196

| | |
|------------------------|-------------------|
| Volume: | 100,000 ml |
| Tolerance: | $\pm 0,1000$ ml |
| Average value: | 99,9969 ml |
| Deviation: | 0,0143 ml |
| Precision (CV): | 0,0143 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

18.11.2020

Balance:

Sartorius -
LE244S-OCE

DE-M