


# Certificate of performance

Article number 3 662 010 USP  
Batch 05-17  
Serial number  
Measuringnr./Batch size 10076 / 499  
Balance Sartorius - LE244S-0CE  
Operator 44  
Date 02.02.2017

## Volumetric flask 100 ml NS 12 class A KB

Volume 100,0000ml  
Tolerance 0,0800ml  
Average volume 100,0135 ml  
Deviation 0,0083 ml  
Number of measurings 20



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
| info@witeg.de

