

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 010

Volumetric flask, 100 ml, NS 12, A, conf. cert.


Batch: 39-24

Serial:

Measuring Nr. 69417

Batch size: 3198

Volume:	100,000 ml
Tolerance:	$\pm 0,1000$ ml
Average value:	100,0107 ml
Deviation:	0,0158 ml
Precision (CV):	0,0158 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	08.10.2024
Balance:	Sartorius - MSE224S-0CE -DU

DE-M