


# Certificate of performance

Article number 3 662 011 BL  
Batch 04-17  
Serial number  
Measuringnr./Batch size 10075 / 1200  
Balance Sartorius - LE244S-0CE  
Operator 4  
Date 02.02.2017

## Volumetric flask 100 ml NS 14 class A KB

Volume 100,0000ml  
Tolerance 0,1000ml  
Average volume 100,0097 ml  
Deviation 0,0196 ml  
Number of measurings 35



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
| info@witeg.de



ISO 9001  
CERTIFIED

