


Certificate of performance

Article number 3 662 012 BL
Batch 02-17
Serial number
Measuringnr./Batch size 10070 / 1191
Balance Sartorius - LE244S-0CE
Operator 4
Date 31.01.2017

Volumetric flask 10 ml NS 10 Class A KB

Volume 10,0000ml
Tolerance 0,0400ml
Average volume 10,0000 ml
Deviation 0,0073 ml
Number of measurings 35



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
| info@witeg.de



ISO 9001
CERTIFIED

