

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 012 W

Vol. flasks, 10 ml, NS 10, white grad., A, conf. cert.


Batch: 16-20

Serial:

Measuring Nr. 37544

Batch size: 278

| | |
|------------------------|------------------|
| Volume: | 10,000 ml |
| Tolerance: | $\pm 0,4000$ ml |
| Average value: | 9,9909 ml |
| Deviation: | 0,0081 ml |
| Precision (CV): | 0,0808 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

| | |
|------------------|---------------------------|
| Operator: | 7 |
| Date: | 29.04.2020 |
| Balance: | Sartorius - LE244S-OCE |

DE-M