

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 012 W

Vol. flasks, 10 ml, NS 10, white grad., A, conf. cert.

Batch: 42-22


Serial:

Measuring Nr. 60348

Batch size:

150

Volume:	10,000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	9,9989 ml
Deviation:	0,0072 ml
Precision (CV):	0,0723 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	18.11.2022
Balance:	Sartorius - MSE224S-0CE -DU

DE-M