

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 012

**Volumetric flasks, 10 ml, NS 10, A, conf. cert.**


**Batch:** 17-22

**Serial:**

**Measuring Nr.** 55247

**Batch size:** 1196

|                        |                  |
|------------------------|------------------|
| <b>Volume:</b>         | <b>10,000 ml</b> |
| <b>Tolerance:</b>      | $\pm 0,0400$ ml  |
| <b>Average value:</b>  | 10,0068 ml       |
| <b>Deviation:</b>      | 0,0036 ml        |
| <b>Precision (CV):</b> | 0,0361 %         |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 7

**Date:** 06.05.2022

**Balance:** Sartorius -  
LE244S-OCE

DE-M