


Certificate of performance

Article number 3 662 014 BL
Batch 03-17
Serial number
Measuringnr./Batch size 10061 / 1170
Balance Sartorius - LE244S-0CE
Operator 4
Date 23.01.2017

Volumetric flask 25 ml Class A NS 12 KB without plug

Volume 25,0000ml
Tolerance 0,0600ml
Average volume 24,9922 ml
Deviation 0,0122 ml
Number of measurings 35



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
| info@witeg.de

