


# Certificate of performance

Article number 3 662 014  
Batch 07-17  
Serial number  
Measuringnr./Batch size 10107 / 1193  
Balance Sartorius - LE244S-0CE  
Operator 44  
Date 14.02.2017

## Volumetric flask 25 ml NS 12 class A KB

Volume 25,0000ml  
Tolerance 0,0600ml  
Average volume 24,9928 ml  
Deviation 0,0061 ml  
Number of measurings 35



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
| info@witeg.de



ISO 9001  
CERTIFIED

