


Certificate of performance

Article number 3 662 016
Batch 03-17
Serial number
Measuringnr./Batch size 10068 / 279
Balance Sartorius - LE244S-0CE
Operator 4
Date 30.01.2017

Volumetric flask 15 ml NS 10 Class A KB

Volume 15,0000ml
Tolerance 0,0400ml
Average volume 14,9931 ml
Deviation 0,0046 ml
Number of measurings 15



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
| info@witeg.de



ISO 9001
CERTIFIED

