

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 016

Volumetric flasks, 15 ml, NS 10, A, conf. cert.


Batch: 43-18

Serial:

Measuring Nr. 25451

Batch size: 277

Volume:	15,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	14,9939 ml
Deviation:	0,0041 ml
Precision (CV):	0,0274 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 9

Date: 07.11.2018

Balance: Sartorius -
LE244S-OCE

DE-M