

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 020 BL

Vol. flasks, 200 ml, NS 14, blue grad., A, conf. cert.

Batch: 48-18


Serial:

Measuring Nr. 26514

Batch size:

1194

Volume:	200,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	200,0209 ml
Deviation:	0,0381 ml
Precision (CV):	0,0191 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	14.12.2018
Balance:	Sartorius - LA 1200 S

DE-M