

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 025

Volumetric flask, 250 ml, NS 14, A, conf. cert.

Batch: 48-19


Serial:

Measuring Nr. 34894

Batch size:

1200

Volume:	250,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	250,0160 ml
Deviation:	0,0333 ml
Precision (CV):	0,0133 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

05.12.2019

Balance:

Sartorius - LA
1200 S

DE-M