

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 050 BL

Vol. flask, 500 ml, NS 19, blue grad., A, conf. cert.

Batch: 31-22


Serial:

Measuring Nr. 57853

Batch size:

1184

Volume:	500,000 ml
Tolerance:	$\pm 0,2500$ ml
Average value:	500,0437 ml
Deviation:	0,0388 ml
Precision (CV):	0,0078 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

15.08.2022

Balance:

Sartorius - LA
1200 S

DE-M