

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 050 USP

**Volumetric flasks, 500 ml, NS 19, USP, A, conf. cert.**

**Batch:** 36-24


**Serial:**

**Measuring Nr.** 69254

**Batch size:**

279

<b>Volume:</b>	500,000 ml
<b>Tolerance:</b>	$\pm 0,2000$ ml
<b>Average value:</b>	500,0130 ml
<b>Deviation:</b>	0,0336 ml
<b>Precision (CV):</b>	0,0067 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	25.09.2024
<b>Balance:</b>	Sartorius - LA 1200 S

DE-M