

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 060

Volumetric flask, 600 ml, NS 24, A, conf. cert.


Batch: 23-20

Serial:

Measuring Nr. 38567

Batch size: 50

Volume:	600,000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	600,0045 ml
Deviation:	0,0345 ml
Precision (CV):	0,0057 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 10.06.2020

Balance: Sartorius - LC
6200 S

DE-M