

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 090

Volumetric flask, 900 ml, NS 24, A, conf. cert.

Batch: 40-22


Serial:

Measuring Nr. 60063

Batch size:

14

Volume:	900,000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	900,1742 ml
Deviation:	0,0654 ml
Precision (CV):	0,0073 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	02.11.2022
Balance:	Sartorius - LA 1200 S

DE-M