

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 100 USP

**Volumetric flask, 1000 ml, NS 24, USP, A, conf. cert.**


**Batch:** 27-19

**Serial:**

**Measuring Nr.** 31250

**Batch size:** 150

<b>Volume:</b>	1000,0000 ml
<b>Tolerance:</b>	± 0,3000 ml
<b>Average value:</b>	1000,1222 ml
<b>Deviation:</b>	0,0386 ml
<b>Precision (CV):</b>	0,0039 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	08.07.2019
<b>Balance:</b>	Sartorius - LC 6200 S

DE-M