

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 100

Volumetric flasks, 1000 ml, NS 24, A, conf. cert.

Batch: 36-17


Serial:

Measuring Nr. 14225

Batch size:

500

Volume:	1000,0000 ml
Tolerance:	$\pm 0,4000$ ml
Average value:	1000,0954 ml
Deviation:	0,0730 ml
Precision (CV):	0,0073 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

20.09.2017

Balance:

Sartorius - LC
6200 S

DE-M