

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 150

**Volumetric flasks, 150 ml, NS 14, A, conf. cert.**

**Batch:** 26-18


**Serial:**

**Measuring Nr.** 22021

**Batch size:**

320

<b>Volume:</b>	<b>150,000 ml</b>
<b>Tolerance:</b>	$\pm 0,1500$ ml
<b>Average value:</b>	149,9750 ml
<b>Deviation:</b>	0,0190 ml
<b>Precision (CV):</b>	0,0127 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	44
<b>Date:</b>	06.07.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M