

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 200 BL

**Vol. flask, 2000 ml, NS 29, blue grad., A, conf. cert.**


**Batch:** 49-19

**Serial:**

**Measuring Nr.** 35101

**Batch size:** 280

|                        |              |
|------------------------|--------------|
| <b>Volume:</b>         | 2000,0000 ml |
| <b>Tolerance:</b>      | ± 0,6000 ml  |
| <b>Average value:</b>  | 2000,1809 ml |
| <b>Deviation:</b>      | 0,1735 ml    |
| <b>Precision (CV):</b> | 0,0087 %     |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

|                  |                          |
|------------------|--------------------------|
| <b>Operator:</b> | 4                        |
| <b>Date:</b>     | 10.12.2019               |
| <b>Balance:</b>  | Sartorius - LC<br>6200 S |

