

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 262 BL

Volumetric flask, 262 ml, NS 14, blue, A, conf. cert.


Batch: 19-23

Serial:

Measuring Nr. 63038

Batch size: 25

Volume:	262,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	262,0212 ml
Deviation:	0,0235 ml
Precision (CV):	0,0090 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	12.05.2023
Balance:	Sartorius - LA 1200 S

DE-M