

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 300 BL

**Vol. flask, 3000 ml, NS 29, blue grad., A, conf. cert.**


**Batch:** 29-20

**Serial:**

**Measuring Nr.** 39265

**Batch size:** 25

<b>Volume:</b>	<b>3000,0000 ml</b>
<b>Tolerance:</b>	$\pm 0,8000$ ml
<b>Average value:</b>	2999,9460 ml
<b>Deviation:</b>	0,2100 ml
<b>Precision (CV):</b>	0,0070 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	28.07.2020
<b>Balance:</b>	Sartorius - LC 6200 S

DE-M