

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 300 BL

**Vol. flask, 3000 ml, NS 29, blue grad., A, conf. cert.**

**Batch:** 35-23


**Serial:**

**Measuring Nr.** 64494

**Batch size:**

14

<b>Volume:</b>	3000,0000 ml
<b>Tolerance:</b>	± 0,8000 ml
<b>Average value:</b>	3000,2269 ml
<b>Deviation:</b>	0,1432 ml
<b>Precision (CV):</b>	0,0048 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

26.09.2023

**Balance:**

Sartorius -  
MSE14202S-  
OCE-D0

DE-M