

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 662 500 BL

Vol. flask, 5000 ml, NS 34, blue grad., A, conf. cert.

Batch: 10-18


Serial:

Measuring Nr. 25118

Batch size:

1

Volume:	5000,0000 ml
Tolerance:	$\pm 1,2000$ ml
Measured value:	5000,0690 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	23.10.2018
Balance:	Sartorius - MSE14202S- OCE-D0

DE-M