

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 662 900 BL

**Vol. flask, 10000 ml, NS 45, blue grad., A, conf. cert.**

**Batch:** 03-22


**Serial:**

**Measuring Nr.** 51470

**Batch size:**

6

<b>Volume:</b>	10000,0000 ml
<b>Tolerance:</b>	$\pm 2,0000$ ml
<b>Average value:</b>	10000,6992 ml
<b>Deviation:</b>	0,3665 ml
<b>Precision (CV):</b>	0,0037 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

21.01.2022

**Balance:**

Sartorius -  
MSE14202S-  
OCE-DO

DE-M