

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 669 001 10 BR

Vol. flasks, 2 ml, NS 10, amber glass, A, conf. cert.


Batch: 27-23

Serial:

Measuring Nr. 63687

Batch size: 149

Volume:	2,0000 ml
Tolerance:	± 0,0400 ml
Average value:	1,9943 ml
Deviation:	0,0046 ml
Precision (CV):	0,2291 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.07.2023
Balance:	Sartorius - LE244S-OCE

DE-M