


Certificate of performance

Article number 3 669 001 10
Batch 48-16
Serial number
Measuringnr./Batch size 10014 / 150
Balance Sartorius - LE244S-0CE
Operator 4
Date 12.12.2016

Volumetric flask 2 ml NS 10 Cl A

Volume 2,0000ml
Tolerance 0,0400ml
Measuring volume 2,0010 ml
Deviation 0,0488 %
Number of measurings 10



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
| info@witeg.de

