

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 669 005 BL

Vol. flask, 10 ml, blue grad., trapezoidal, A, conf. cert.


Batch: 28-21

Serial:

Measuring Nr. 47942

Batch size: 98

Volume:	10,000 ml
Tolerance:	$\pm 0,0250$ ml
Average value:	10,0014 ml
Deviation:	0,0025 ml
Precision (CV):	0,0253 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 16.07.2021

Balance: Sartorius -
LE244S-OCE

DE-M