

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 001

Vol. flask, 10 ml, NS 7, amber glass, A, conf. cert.


Batch: 14-23

Serial:

Measuring Nr. 62582

Batch size: 817

Volume:	10,0000 ml
Tolerance:	$\pm 0,0250$ ml
Average value:	10,0007 ml
Deviation:	0,0030 ml
Precision (CV):	0,0300 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 18.04.2023

Balance: Sartorius -
LE244S-OCE

DE-M