

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 670 002 USP

**Vol.flask, 20 ml, USP, NS 10, amber glass, A, conf.cert**

**Batch:** 36-21


**Serial:**

**Measuring Nr.** 48558

**Batch size:**

150

<b>Volume:</b>	20,000 ml
<b>Tolerance:</b>	± 0,0300 ml
<b>Average value:</b>	20,0002 ml
<b>Deviation:</b>	0,0072 ml
<b>Precision (CV):</b>	0,0362 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

31.08.2021

**Balance:**

Sartorius -  
LE244S-OCE

DE-M